SOFT 2018



Contribution ID: 645

Type: not specified

P2.073 Study on the effects of the Signal to Noise ratio on the error in counting of the ITER Radial Neutron Camera (RNC)

Tuesday, 18 September 2018 11:00 (2 hours)

RNC main purpose is to measure the uncollided 14 MeV and 2.5 MeV neutrons from deuterium-tritium (DT) and deuterium-deuterium (DD) fusion reactions through an array of detectors located in collimated lines of sight (LOS) viewing the plasma through the ITER Equatorial Port Plug #1. The measures will be used to calculate the neutron emission profile (neutron emissivity)

The environment where the RNC is be located will be very noisy, affecting the precision of the measure of the neutron and gamma particles count rates.

The paper present a study focusing on finding the maximum allowed S/N ratio, mantaining the counting error below a certain threshold (10%). Experiments will be carried out with the CAEN 5800 detector emulator, which is able to produce a particle spectra similar to the one expected in ITER.

Acknowledgments: The work leading to this publication has been funded partially by Fusion for Energy under the Contract F4E-FPA-327. This publication reflects the views only of the author, and Fusion for Energy cannot be held responsible for any use which may be made of the information contained therein.

Presenter: RIVA, Marco (FSN ENEA)

Session Classification: P2